



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ BIVALIRUDIN CONTROL
Product Code : 225701
Lot No. : FD179817 / FD180217
Expiration date : 2027-03-08

Components	Quantity	Lot	Expiration Date
C1	3 vials	FD1798de17 / FD180217	2027-03-08
C2	3 vials	FD1798df17 / FD180217	2027-03-08

Analytical Data	Specification	Result
Concentration	C1 : 1.25 -1.75 µg/ml C2 : 2.50 - 3.30 µg/ml	1.69 µg/ml 3.13 µg/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval
 Date of Approval : 03 OCT. 2024
 S. LECOURT
 Quality Control Manager :

