



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description :</b>	<b>BIOPHEN APIXABAN CALIBRATOR</b>
<b>Product Code :</b>	<b>226201</b>
<b>Lot No. :</b>	<b>FD17851F</b>
<b>Expiration date :</b>	<b>2027-03-06</b>

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD1785DO1F	2027-03-06
CAL 2	4 vials	FD1785DP1F	2027-03-06
CAL 3	4 vials	FD1785DQ1F	2027-03-06

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 50$ ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 318 ng/mL 552 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

**Quality Control Approval**

Date of Approval : **01 OCT. 2024**  
**S. LECOURT**

Quality Control Manager :

