



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor XI Deficient Plasma
Product Code : DP070A/K
Lot No. : FD19871E
Expiration date : 2027-03-29

Components	Quantity	Lot	Expiration Date
Factor XI Deficient Plasma	1 vial (DP070A) 6 vials (DP070K)	FD1987JX1E	2027-03-29

Analytical Data	Specification	Result
Calibration curve	CT (0%) >150 sec CT (100%) <100 sec	181sec 60sec
Concentration FXI	< 1%	<1%
Concentration FV	> 50 %	93%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

23 OCT. 2024
S. LECOURT

Quality Control Manager :