



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN AT

Item Code : 120002

Lot No. : FD2121NC / FD2137NF

Expiration date : 2027-04-23

| Components | Quantity | Lot | Expiration Date |
|----------------------|----------|----------------------|-----------------|
| R1 : Latex reagent | 2 vials | FD2121WNC / FD2137NF | 2027-04-23 |
| R2 : Reaction Buffer | 2 vials | FD2121XNC / FD2137NF | 2027-04-23 |

| Analytical Data | Specification | Result |
|---|---|-----------------------|
| Calibration curve : reactivity, slope, linearity and accuracy | A620 (0-150%) ≥ 0.40 R ² ≥ 0.98 Controls measured within acceptance ranges | 0.805 PASS PASS |
| Detection threshold | ≤ 10% | <4% |
| Hook effect | >250% | PASS |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval : **20 NOV. 2024**

Quality Control Manager : **S. LECOURT**

