



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

**CERTIFICATE OF ANALYSIS**

**Product Description:** BIOPHEN LMWH CONTROL C3  
**Product Code:** 223801  
**Lot No.:** FD2157NL / FD2181NR  
**Expiration date:** 2027-04-19

Components	Quantity	Lot	Expiration Date
C3	12 vials	FD2157BONL / FD2181NR	2027-04-19

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.77 IU/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: **Quality Control Approval** Quality Control Manager:

16 DEC. 2024

S. LECOURT

