



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ARIXTRA CONTROL PLASMA  
**Product Code :** 224001  
**Lot No. :** FD21091C  
**Expiration date :** 2027-04-13

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD2109CR1C	2027-04-13
C2	6 vials	FD2109CS1C	2027-04-13

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/ml C2: 1.05 -1.35 µg/ml	C1 : 0.43 µg/ml C2 : 1.13 µg/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**  
**05 NOV. 2024**  
**S. LECOURT**  
Quality Control Manager :