



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN™ BIVALIRUDIN CALIBRATOR  
**Product Code :** 226701  
**Lot No. :** FD21021A  
**Expiration date :** 2027-04-13

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FD21021A	2027-04-13
CAL 2	2 vials	FD21021A	2027-04-13
CAL 3	2 vials	FD21021A	2027-04-13
CAL 4	2 vials	FD21021A	2027-04-13
CAL 5	2 vials	FD21021A	2027-04-13

Analytical Data	Specification	Result
Concentration	CAL 1 : < 0.25 µg/mL CAL 2 : 0.8 -1.2 µg/mL CAL 3 : 1.6 -2.4 µg/mL CAL 4 : 3.0 - 4.0 µg/mL CAL 5 : 4.0 - 5.0 µg/mL	0 µg/mL 1.12 µg/mL 2.17 µg/mL 3.60 µg/mL 4.41 µg/mL
Accuracy of calibration with values assigned	Measured values for controls in acceptance ranges	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

### Quality Control Approval

Date of Approval :

14 NOV. 2024

Quality Control Manager :

S. LECOURT

