



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor VII Deficient Plasma
Product Code : DP030A/K
Lot No. : FD20321N / FD208113
Expiration date : 2027-04-06

Components	Quantity	Lot	Expiration Date
Factor VII Deficient Plasma	1 vial (DP030A) 6 vials (DP030K)	FD2032JR1N / FD208113	2027-04-06

Analytical Data	Specification	Result
Calibration curve	CT (0%) >100 SEC CT (100%) <40 SEC Thromboplastin: STAGO	138sec 27sec
Concentration FVII	< 1%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

11 MARS 2025

Sophie LECOURT
Deputy Director Quality Control

Quality Control Manager :