



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ORGARAN CALIBRATOR  
**Product Code :** 222201  
**Lot No. :** FD24921A / FD252513  
**Expiration date :** 2027-05-29

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD2492CG1A / FD252513	2027-05-29
CAL 2	4 vials	FD2492CH1A / FD252513	2027-05-29
CAL 3	4 vials	FD2492CI1A / FD252513	2027-05-29
CAL 4	4 vials	FD2492CJ1A / FD252513	2027-05-29
CAL 5	4 vials	FD2492CK1A / FD252513	2027-05-29

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05$ U/ml CAL 2 : 0.30 -0.50 U/ml CAL 3 : 0.70 -0.90 U/ml CAL 4 : 1.10 -1.30 U/ml CAL 5 : 1.45 -1.75 U/ml	0.00 U/mL 0.38 U/mL 0.81 U/mL 1.19 U/mL 1.57 U/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

### Quality Control Approval

Date of Approval :

**19 DEC. 2024**  
**S. LECOURT**

Quality Control Manager :