



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ARIXTRA CALIBRATOR  
**Product Code :** 222501  
**Lot No. :** FD24311A / FD24531B  
**Expiration date :** 2027-05-22

| Components | Quantity | Lot                   | Expiration Date |
|------------|----------|-----------------------|-----------------|
| CAL 1      | 3 vials  | FD2431CN1A / FD24531B | 2027-05-22      |
| CAL 2      | 3 vials  | FD2431CO1A / FD24531B | 2027-05-22      |
| CAL 3      | 3 vials  | FD2431CP1A / FD24531B | 2027-05-22      |
| CAL 4      | 3 vials  | FD2431CQ1A / FD24531B | 2027-05-22      |

| Analytical Data  | Specification  | Result   |
|--|--|--|
| Concentration  | CAL 1 : $\leq 0.05 \mu\text{g/ml}$<br>CAL 2 : $0.40 - 0.60 \mu\text{g/ml}$<br>CAL 3 : $0.90 - 1.10 \mu\text{g/ml}$<br>CAL 4 : $1.35 - 1.60 \mu\text{g/ml}$ | 0.00 $\mu\text{g/ml}$<br>0.49 $\mu\text{g/ml}$<br>0.96 $\mu\text{g/ml}$<br>1.42 $\mu\text{g/ml}$ |
| Linearity and accuracy of calibration with values assigned | $R^2 \geq 0.98$<br>Measured values for controls in acceptance ranges   | PASS<br>PASS   |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

16 DEC. 2024

S. LECOURT