



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT THROMBIN (BOVINE)
Product Code : CK011K
Lot No. : FD2381PP / FD2407PR
Expiration date : 2027-05-14

Components	Quantity	Lot	Expiration Date
Bovine Thrombin	6 vials	FD2381YPP / FD2407PR	2027-05-14

Analytical Data	Specification	Result
Thrombin Time on normal human citrated plasma	15-25 sec	17.7 sec on STAR 16.5 sec on CS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

12 DEC. 2024

S. LECOURT