



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ Dabigatran Control Plasma
Product Code : 224701
Lot No. : FE00851G / FE014013
Expiration date : 2027-07-03

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE0085AV1G	2027-07-03
C2	6 vials	FE0085AW1G	2027-07-03

Analytical Data	Specification	Result
Concentration	C1 : 50 -200 ng/ml C2 : 200 -300 ng/ml	105 ng/mL 254 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
30 JAN. 2025
S. LECOURT
Quality Control Manager :