



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FE2078 / FE2079
Expiration date : 2027-08-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE0457BX39 / FE04693B	2027-08-09
C2	6 vials	FE0458BY3A / FE04703C	2027-08-09

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.20 IU/mL 0.57 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

06 MARS 2025
S. LECOURT

Quality Control Manager :